

RESEARCH ARTICLE

Introducing automation to the molecular diagnosis of *Trypanosoma cruzi* infection: A comparative study of sample treatments, DNA extraction methods and real-time PCR assays

Alba Abras^{1,2,3*}, Cristina Ballart^{1,2*}, Teresa Llovet^{4,5}, Carme Roig⁴, Cristina Gutiérrez⁴, Silvia Tebar^{1,2}, Pere Berenguer⁴, María-Jesús Pinazo², Elizabeth Posada², Joaquim Gascón², Alejandro G. Schijman⁶, Montserrat Gállego^{1,2*}, Carmen Muñoz^{4,5,7*}

1 Secció de Parasitologia, Departament de Biologia, Sanitat i Medi Ambient, Facultat de Farmàcia i Ciències de l'Alimentació, Universitat de Barcelona, Barcelona, Spain, **2** ISGlobal, Barcelona Centre for International Health Research (CRESIB), Hospital Clínic–Universitat de Barcelona, Barcelona, Spain, **3** Laboratori d'Ictiologia Genètica, Departament de Biologia, Universitat de Girona, Girona, Spain, **4** Servei de Microbiologia, Hospital de la Santa Creu i Sant Pau, Barcelona, Spain, **5** Departament de Genètica i Microbiologia, Universitat Autònoma de Barcelona, Cerdanyola del Vallès, Spain, **6** Laboratorio de Biología Molecular de la Enfermedad de Chagas (LaBMECh), Instituto de Investigaciones en Ingeniería Genética y Biología Molecular “Dr. Héctor N. Torres” (INGEBI-CONICET), Buenos Aires, Argentina, **7** Institut d'Investigació Biomèdica Sant Pau (IIB Sant Pau), Hospital de la Santa Creu i Sant Pau, Barcelona, Spain

* These authors contributed equally to this work.

* mgallego@ub.edu (MG); cmunoz@santpau.cat (CM)



OPEN ACCESS

Citation: Abras A, Ballart C, Llovet T, Roig C, Gutiérrez C, Tebar S, et al. (2018) Introducing automation to the molecular diagnosis of *Trypanosoma cruzi* infection: A comparative study of sample treatments, DNA extraction methods and real-time PCR assays. PLoS ONE 13(4): e0195738. <https://doi.org/10.1371/journal.pone.0195738>

Editor: Ulrich Melcher, Oklahoma State University, UNITED STATES

Received: December 27, 2017

Accepted: March 28, 2018

Published: April 17, 2018

Copyright: © 2018 Abras et al. This is an open access article distributed under the terms of the [Creative Commons Attribution License](https://creativecommons.org/licenses/by/4.0/), which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited.

Data Availability Statement: The data underlying this study cannot be made publicly available as protocols approved by the Ethics Committees of the participating research centers (Clinical Research Ethics Committee of Hospital de la Santa Creu i Sant Pau, Ethics Review Committee of Hospital Clínic and Research Ethics Committee of Universitat de Barcelona) explicitly indicated that patients' data would be anonymized and confidentiality preserved. Public availability would compromise patient privacy and breach

Abstract

Background

Polymerase chain reaction (PCR) has become a useful tool for the diagnosis of *Trypanosoma cruzi* infection. The development of automated DNA extraction methodologies and PCR systems is an important step toward the standardization of protocols in routine diagnosis. To date, there are only two commercially available Real-Time PCR assays for the routine laboratory detection of *T. cruzi* DNA in clinical samples: TCRUZIDNA.CE (Diagnostic Bioprobes Srl) and RealCycler CHAG (Progenie Molecular). Our aim was to evaluate the RealCycler CHAG assay taking into account the whole process.

Methodology/Principal findings

We assessed the usefulness of an automated DNA extraction system based on magnetic particles (EZ1 Virus Mini Kit v2.0, Qiagen) combined with a commercially available Real-Time PCR assay targeting satellite DNA (SatDNA) of *T. cruzi* (RealCycler CHAG), a methodology used for routine diagnosis in our hospital. It was compared with a well-known strategy combining a commercial DNA isolation kit based on silica columns (High Pure PCR Template Preparation Kit, Roche Diagnostics) with an *in-house* Real-Time PCR targeting SatDNA. The results of the two methodologies were in almost perfect agreement, indicating they can be used interchangeably. However, when variations in protocol factors were applied (sample treatment, extraction method and Real-Time PCR), the results were less

compliance with protocols approved by the ethics committees of the participant centers. Nevertheless, researchers who need to access the data may request the information from Dr. Ferran Navarro (fnavarro@santpau.cat), the person in charge of research at Microbiology Service of Hospital de la Santa Creu i Sant Pau.

Funding: This study was partially supported by Departament d'Universitats, Recerca i Societat de la Informació de la Generalitat de Catalunya (grant 2014SGR026), Institut d'Investigació Biomèdica Sant Pau (IIB Sant Pau) and by Plan Nacional de I +D+i and Instituto de Salud Carlos III, Subdirección General de Redes y Centros de Investigación Cooperativa, Ministerio de Economía y Competitividad, Spanish Network for Research in Infectious Diseases (grant REIPI RD12/0015); cofinanced by European Development Regional Fund (ERDF) "A way to achieve Europe". A.A., C.B., S.T., M.-J.P., E.P., J.G., and M.G. belong to RICET, a Tropical Disease Cooperative Research Network in Spain (grant RD12/0018/0010). ISGlobal is a member of the CERCA Programme, Generalitat de Catalunya. The supporters had no role in study design, data collection and analysis, decision to publish, or preparation of the manuscript.

Competing interests: The authors have declared that no competing interests exist.

convincing. A comprehensive fine-tuning of the whole procedure is the key to successful results. Guanidine EDTA-blood (GEB) samples are not suitable for DNA extraction based on magnetic particles due to inhibition, at least when samples are not processed immediately.

Conclusions/Significance

This is the first study to evaluate the RealCycler CHAG assay taking into account the overall process, including three variables (sample treatment, extraction method and Real-Time PCR). Our findings may contribute to the harmonization of protocols between laboratories and to a wider application of Real-Time PCR in molecular diagnostic laboratories associated with health centers.

Introduction

Chagas disease, a parasitic infection caused by the protozoan *Trypanosoma cruzi*, is endemic in 21 countries of Latin America, with approximately six million people affected [1]. Migratory flows have expanded Chagas disease worldwide, especially since the beginning of 2000, and the disease has emerged in non-endemic countries of North America, Europe and the Western Pacific Region [2,3].

In endemic settings the parasite is mainly transmitted by blood-sucking triatomine bugs [4,5], whereas in areas without vector-borne exposure the risk of developing *T. cruzi* infection arises from congenital transmission, blood transfusion, organ transplant, and laboratory accidents [6,7]. The disease has two stages: acute and chronic. Although both can be asymptomatic, the chronic phase is usually associated with cardiac and gastrointestinal disorders as well as low and intermittent parasitemia [8,9].

Several studies carried out in Spain and other parts of Europe have reported that screening for Chagas disease and early diagnosis is cost-effective [10–12]. However, a large number of patients in both endemic and non-endemic countries are still diagnosed late or not at all. Therefore, there is an urgent need to establish an efficient diagnostic strategy to deal with *T. cruzi* infection [13].

The serological diagnosis for Chagas disease is often difficult to interpret and although the use of a single test has been recently proposed [14], there is still no reference standard [15]. Serological methods are currently widely used, especially in the chronic phase, and dozens of tests are commercially available [16–19]. Nevertheless, serological diagnosis has certain disadvantages: the persistence of positive results in chronically infected patients for years after treatment [20], the possibility of cross-reactions with other trypanosomatids like *Trypanosoma rangeli* or *Leishmania* spp. [21], and the transmission of passive antibodies from mother to newborn [22–24].

Molecular biology techniques, especially the polymerase chain reaction (PCR), have been proposed as useful tools for the diagnosis of *T. cruzi* infection [25,26]. Unlike the serology test, a positive PCR result confirms the presence of the parasite DNA. Moreover, the high sensitivity of PCR in comparison with classical parasitological techniques is particularly useful for an early diagnosis of congenital cases [27–29].

Molecular detection of *T. cruzi* is also important in the chronic phase, as it can detect the therapeutic failure of anti-parasitic treatments and parasite reactivation in patients with

immunosuppression [30,31]. Furthermore, the PCR variant known as quantitative Real-Time PCR also quantifies the amplification product using intercalating dyes or labeled probes and standard curves of known parasite concentration [32]. The main problem with this approach stems from the lack of consensus among laboratories on the PCR strategies used.

The recent development of automated DNA extraction methodologies coupled to PCR systems is an important step toward protocol standardization, while also preventing contamination of samples and reagents due to human manipulation. Automated nucleic acid extraction systems are usually based on magnetic separation, a time-saving technology in comparison with silica column-based DNA extractions [33,34]. One of the most applied molecular protocols is the amplification of satellite DNA (SatDNA) of *T. cruzi* [35,36]. SatDNA is the most abundant repetitive sequence in the parasite nuclear genome and is composed of about 10^5 copies of a 195 nucleotide repeat [37,38].

To date, there are only two commercially available Real-Time PCR assays for the diagnosis of *T. cruzi* infection: TCRUZIDNA.CE (Diagnostic Bioprobes Srl, Sesto San Giovanni, Italy) and RealCycler CHAG (Progenie Molecular, Valencia, Spain). Both amplify the SatDNA sequence of *T. cruzi*. Seiringer et al. [39] recently evaluated TCRUZIDNA.CE but RealCycler CHAG has not been assessed until now, and no studies have previously evaluated these *T. cruzi* Real-Time PCR assays combined with different DNA extraction methods.

The aim of the present study was to assess the usefulness of an automated DNA extraction system based on magnetic particles (EZ1 Virus Mini Kit v2.0, Qiagen, Hilden, Germany) combined with a commercially available Real-Time PCR assay that targets the SatDNA of *T. cruzi* (RealCycler CHAG), a methodology routinely used for *T. cruzi*-infection diagnosis in the Hospital de la Santa Creu i Sant Pau of Barcelona (Spain). This technique was compared with a well-known and widely used strategy combining a commercial DNA isolation kit based on silica columns (High Pure PCR Template Preparation Kit, Roche Diagnostics, Mannheim, Germany) with an *in-house* Real-Time PCR that also targets the parasite SatDNA sequence [36].

Materials and methods

Ethics statement

Approval was obtained from the participating centers, the Clinical Research Ethics Committee of the Hospital de la Santa Creu i Sant Pau, the Ethics Review Committee of the Hospital Clínic and the Research Ethics Committee of the Universitat de Barcelona. All samples were anonymized before being evaluated and included in the study.

Samples and population

A total of 123 blood samples were used in this study. Samples were collected from Hospital de la Santa Creu i Sant Pau and Hospital Clínic of Barcelona (Spain) during the period from January 2013 to March 2017. EDTA-blood (EB) and EDTA-blood mixed with an equal volume of guanidine hydrochloride solution 6 M (GEB) were obtained. Since the samples in the study were obtained retrospectively, in some cases only one of the two options, EB or GEB, was available. EB samples were stored at -40°C and GEB samples at 4°C until the subsequent analysis. Sample distribution was as follows: 19 from non-infected newborns from chagasic mothers, 5 from non-chagasic adults from endemic countries, 12 from non-chagasic individuals from non-endemic countries, 65 from non-treated chronic chagasic patients, and 22 seronegative GEB samples experimentally spiked with cultured epimastigotes of *T. cruzi* stocks as well as non-spiked GEB samples. Sample results were defined in the following way: non-infected newborns from chagasic mothers were followed up by serology until their negativization within

the first year of life, non-chagasic adults were those with negative serology, and chronic chagasic patients were serologically positive by at least two different immunological techniques.

DNA extraction procedures

Two different methodologies were used for the DNA extraction.

(i) Silica gel columns—High Pure PCR Template Preparation Kit (Roche Diagnostics). The protocol is available online at <https://lifescience.roche.com>. DNA extraction was performed according to the manufacturer's instructions. DNA was extracted from 200 μ L of EB or 300 μ L of GEB and eluted in 100 μ L of the elution buffer supplied with the kit, as previously described [40]. The extracted DNA was stored at -40°C until its analysis.

(ii) Magnetic particles—EZ1 Virus Mini Kit v2.0 (Qiagen). The protocol is available online at <https://www.qiagen.com>. DNA extraction was performed according to the manufacturer's instructions using the EZ1 Advanced automated system based on magnetic particles. DNA was extracted from 400 μ L of EB or GEB samples and eluted in 60 μ L of the AVE buffer supplied with the kit. The extracted DNA was stored at -40°C until its analysis. EB samples were pre-treated with an equal volume of the erythrocyte lysis buffer (EL) (Qiagen), and 400 μ L of the mixture was then introduced to the EZ1 device. GEB samples were directly loaded into the system.

For quantification, standard curves were built using non-chagasic human EB and GEB spiked with cultured epimastigotes of the *T. cruzi* Maracay strain (TcI), giving a final concentration of 10^6 parasite equivalents/mL (par. eq./mL). DNA from spiked blood used to construct the standard curve was extracted with the High Pure PCR Template Preparation Kit and EZ1 Virus Mini Kit v2.0, and 1/10 serial dilutions of the DNA extracted in total blood DNA extractions from non-chagasic individuals were carried out in order to obtain a panel of samples ranging from 10^5 to 1 par. eq./mL.

Real-Time PCR methods

Two Real-Time PCR procedures were performed.

(i) *In house* Real-Time PCR targeting the *T. cruzi* SatDNA. Five μ L of the DNA eluates in a final volume of 20 μ L were amplified in triplicate. The final concentrations were: 1 x FastStart Universal Probe Master (Rox) (Roche Diagnostics), 0.75 μ M of each primer *Cruzi* 1 and *Cruzi* 2, 0.25 μ M of the probe *Cruzi* 3, and 0.2 x TaqMan Human RNase P detection reagent (Applied Biosystems, Austin, TX) [36]. The reaction was carried out in an ABI7900 device (Applied Biosystems) and the amplification of the RNase P human gene was included as an internal amplification control (IAC) [36]. Amplification conditions were as follows: one step of 10 min at 95°C , and 40 cycles of 95°C for 15 s, and 58°C for 1 min [40]. A sample was considered valid when the RNase P human gene was efficiently amplified with a cycle threshold value (Ct) ≤ 29 , which was established with the Tukey criterion to detect outliers [41,42]. Samples were classified as inhibited when the IAC gave negative results or the Ct was > 29 . A sample was considered positive when the Ct of the target was ≤ 40 in at least one of the three replicates.

(ii) RealCycler CHAG (Progenie Molecular). The test was carried out according to the manufacturer's instructions using the SmartCycler Automated Real-Time PCR system (Cepheid, Sunnyvale, CA). The kit includes the CHAG AmpliMix, which contains all the reagents necessary for the amplification as well as the IAC. The Real-Time PCR procedure was performed in 25 μ L reaction volume containing 7.5 μ L of eluted DNA and a single replica per sample was amplified. According to the manufacturer's specifications, amplification conditions were as follows: one step of 15 min at 95°C , and 45 cycles of 95°C for 15 s, 60°C for 30 s, and 72°C for

30 s. A sample was considered valid when the IAC was amplified with a $Ct \leq 35$ and inhibited when the criterion was not fulfilled. A sample was considered positive when the Ct of the target was ≤ 40 .

Experimental procedures

Sample treatments (EB and GEB), DNA extraction methods (silica columns and magnetic particles), and Real-Time PCR procedures (*In house* Real-Time PCR and RealCycler CHAG) were combined to create eight different methodological protocols for assessment (Table 1).

The *in house* protocol A was considered the reference standard due to its validation in previous international studies [26,42] and the commercialized protocol H is used for *T. cruzi* routine diagnosis in the Hospital de la Santa Creu i Sant Pau.

Data analysis

Taken as the reference standard, protocol A was compared with all the others (B to H), resulting in seven possible combinations (A-B, A-C, A-D, A-E, A-F, A-G, and A-H). Protocols were also compared by focusing on only one variable of the process: sample treatments, DNA extraction methods or Real-Time PCR procedures. In this way, 12 combinations were created: A-E, B-F, C-G, and D-H comparing sample treatments; A-C, B-D, E-G, and F-H comparing DNA extraction methods; and A-B, C-D, E-F, and G-H comparing Real-Time PCR procedures. For the protocol comparisons, Cohen's kappa coefficient (K), which describes the level of concordance between two tests relating the observed agreement and the agreement expected by chance, was calculated for each protocol combination created. The interpretation of K values was as follows: 0 to 0.20 indicates slight agreement, 0.21 to 0.40 fair agreement, 0.41 to 0.60 moderate agreement, 0.61 to 0.80 substantial agreement, and 0.8 to 1 almost perfect agreement [43]. Calculations were performed with the software EPIDAT 3.1, which is available online at <http://www.sergas.es/Saude-publica>.

Parasitic loads obtained from positive samples in protocols A and H were compared in a Bland-Altman difference plot in order to quantify the agreement between the two methods [44]. The Spearman rank correlation coefficient (r_s), which describes the strength of a monotonic relationship between paired data, was also calculated [45]. Ranging from -1.0 to +1.0, values close to +1.0 indicate positive association whereas values close to -1.0 indicate negative association. The interpretation of the absolute value of r_s is as follows: 0 to 0.19 indicates a very

Table 1. *Trypanosoma cruzi* diagnostic Real-Time PCR results of samples analysed according to the protocols used in the study based on the combination of different sample treatments, DNA extraction methods and Real-Time PCR procedures.

Protocol code	Sample treatment	DNA extraction method	Real-Time PCR	N	Positive samples	Inhibited samples	N excluding inhibition
A (ref.)	GEB	Roche silica columns	<i>In house</i>	123	72	0	123
B	GEB	Roche silica columns	RealCycler	123	58	0	123
C	GEB	Qiagen magnetic particles	<i>In house</i>	62	8	35	27
D	GEB	Qiagen magnetic particles	RealCycler	62	25	17	45
E	EB	Roche silica columns	<i>In house</i>	25	11	0	25
F	EB	Roche silica columns	RealCycler	25	11	0	25
G	EB	Qiagen magnetic particles	<i>In house</i>	64	15	2	62
H	EB	Qiagen magnetic particles	RealCycler	64	15	0	64

Inhibited samples had negative results or a cycle threshold (Ct) value >29 for the internal amplification control in protocols using *in house* Real-Time PCR (A, C, E, and G), and negative results or $Ct >35$ in protocols using RealCycler CHAG (B, D, F, and H).

Ref.: reference standard protocol, EB: EDTA-blood, GEB: Guanidine EDTA-blood, N: number of samples.

<https://doi.org/10.1371/journal.pone.0195738.t001>

weak relationship, 0.20 to 0.39 weak, 0.40 to 0.59 moderate, 0.60 to 0.79 strong, and 0.80 to 1.0 a very strong relationship.

Results

A total of 123 blood samples were analyzed. Since this is a retrospective study, the number of samples analyzed with each protocol is not identical (Table 1).

The results of the comparison between the reference protocol A and the others are summarized in Table 2. K results are shown including and excluding inhibited samples. In protocols using GEB samples a high level of inhibition was found when the DNA extraction method based on magnetic particles was used, as shown in protocols C and D with 56.5% and 27.4% of inhibition, respectively (Table 1). In the other cases, protocols did not present inhibition, with the exception of protocol G (EB, magnetic particles and *in house* Real-Time PCR), which yielded two inhibited samples (3.1%).

The number of discordant results between protocols ranged from 1 to 13. The results obtained with these samples are listed in Table 3.

Protocols A (reference standard) and H (routine diagnosis in our hospital) gave matching qualitative results for the 64 samples analysed ($K = 1$), 15 of which were positive. These protocols also underwent a quantitative evaluation (Table 4). The level of agreement between the positive samples in both methods is represented by a Bland-Altman difference plot (Fig 1). The mean bias was determined as 0.12 Log_{10} par. eq./10 mL, indicating a systematic bias of 0.7-fold parasite equivalents per 10 mL between methods. The bias was in both directions and did not differ in relation to the magnitude of the parasitic loads. However, samples with the highest bias yielded very low parasitic loads (samples S2 and S11 in Table 4). Limits of agreement were also included in the plot expressed as bias $\pm 1.96 \times$ standard deviation (SD). The SD was 0.43, the lower limit of agreement with a 95% confidence interval (CI) was -0.72 Log_{10} par. eq./10 mL and the upper one was 0.96 Log_{10} par. eq./10 mL. The Spearman rank correlation coefficient was $r_s = 0.97$, which indicates very strong and positive correlation between protocols.

Table 2. Cohen's kappa coefficient (K) results for the comparison between the reference protocol A and the others. K values are shown including and excluding inhibited samples.

Protocols	N	K value (95% CI)	Positive samples	Inhibited samples	N excluding inhibition	Discordant samples excluding inhibition	K value excluding inhibited samples (95% CI)
A-B	123	0.79 (0.68–0.9)	59	0	123	13	0.79 (0.68–0.9)
A-C	62	0.17 (0.08–0.26)	8	35	27	8	0.45 (0.18–0.72)
A-D	62	0.29 (0.14–0.44)	24	17	45	10	0.53 (0.3–0.77)
A-E	25	0.92 (0.77–1)	11	0	25	1	0.92 (0.77–1)
A-F	25	0.76 (0.5–1)	10	0	25	3	0.76 (0.5–1)
A-G	64	0.92 (0.81–1)	15	2	62	0	1
A-H	64	1	15	0	64	0	1

N: number of samples assessed in parallel by both protocols compared. Positive samples: number of samples with a positive result for both protocols. Inhibited samples: number of samples with an invalid result for the internal amplification control (IAC) in at least one of the two protocols compared. CI: confidence interval.

<https://doi.org/10.1371/journal.pone.0195738.t002>

Table 3. *Trypanosoma cruzi* diagnostic Real-Time PCR results obtained in the eight protocols assessed for the 22 discordant samples.

Sample ID	Protocols ^a							
	A	B	C	D	E	F	G	H
S1	N	N	N	N	N	P (40) 1/1	N	N
S2	P (32.2) 3/3	P (32.2) 1/1	P (29) 3/3	P (31.9) 1/1	P (31.9) 3/3	N	P (32.9) 3/3	P (35) 1/1
S3	N	N	N	P (40) 1/1	N	N	N	N
S4	P (28.2) 3/3	P (30.5) 1/1	N	P (28.3) 1/1	P (36.2) 3/3	P (39.5) 1/1	P (25.8) 3/3	P (28.5) 1/1
S5	P (29.6) 3/3	P (30.6) 1/1	P (28.5) 3/3	P (31.2) 1/1	N	N	P (26.9) 3/3	P (30.6) 1/1
S6	P (35.1) 1/3	N	I	I	-	-	-	-
S7	P (32.2) 2/3	N	I	P (34.2) 1/1	-	-	-	-
S8	P (35.3) 1/3	N	I	N	-	-	-	-
S9	P (29.9) 3/3	P (34.1) 1/1	N	P (32.2) 1/1	-	-	-	-
S10	P (36.4) 1/3	N	N	N	-	-	-	-
S11	P (35.5) 1/3	N	I	N	-	-	-	-
S12	P (29.4) 3/3	P (31.8) 1/1	N	P (31.2) 1/1	-	-	-	-
S13	P (35.6) 1/3	N	N	N	-	-	-	-
S14	P (35) 1/3	N	I	I	-	-	-	-
S15	P (36.3) 1/3	P (38.4) 1/1	I	N	-	-	-	-
S16	P (35.9) 1/3	N	N	N	-	-	-	-
S17	P (33.4) 2/3	N	I	I	-	-	-	-
S18	P (34.9) 2/3	N	I	N	-	-	-	-
S19	P (33.7) 1/3	N	I	I	-	-	-	-
S20	P (34.6) 1/3	P (39.7) 1/1	I	N	-	-	-	-
S21	P (35.4) 1/3	N	N	I	-	-	-	-
S22	P (35.4) 1/3	N	N	N	-	-	-	-

Protocols A, C, E, and G used the *in house* Real-Time PCR with three replicates per sample. Protocols B, D, F, and H used RealCycler CHAG Real-Time PCR with a single replica amplification per sample.

^aCycle threshold (Ct) results are shown in parentheses. In case of multiple positive replicates, the mean of the Cts obtained is indicated. The number of positive replicates is expressed as a fraction below.

P: positive, N: negative, I: inhibited.

<https://doi.org/10.1371/journal.pone.0195738.t003>

Protocols were also compared taking into account only one variable of the procedure (Tables 5–7). K values and positive and inhibited sample results are shown for the following variables: (i) sample treatment (Table 5); (ii) DNA extraction method (Table 6); and (iii) Real-Time PCR procedures (Table 7). Modifications in one of the three variables resulted in changes in the K values, the number of inhibited samples and the percentage of discordant results obtained.

Table 4. Quantitative evaluation (parasitic load) of the 15 samples positive in protocols A (reference standard) and H (Hospital de la Santa Creu i Sant Pau routine diagnosis).

Sample ID	Protocol A		Protocol H	
	Par. eq./mL	Log ₁₀ Par. eq./10 mL	Par. eq./mL	Log ₁₀ Par. eq./10 mL
S1	0.47	0.67	0.28	0.45
S2	1.26	1.1	0.13	0.11
S3	51.80	2.71	24.79	2.39
S4	4.30	1.63	8.50	1.93
S5	21.90	2.34	21.55	2.33
S6	47.80	2.68	54.09	2.73
S7	12.49	2.1	6.43	1.81
S8	22.39	2.35	40.88	2.61
S9	331.00	3.52	395.66	3.6
S10	0.09	-0.05	0.03	-0.5
S11	0.39	0.59	0.04	-0.4
S12	1.94	1.29	5.21	1.72
S13	3.09	1.49	4.18	1.62
S14	1.14	1.06	0.94	0.97
S15	3.44	1.54	6.43	1.81

Parasitic loads are expressed in parasite equivalents in one mL of blood (Par. eq./mL) and in logarithms of the parasite equivalents in 10 mL of blood (Log₁₀ Par. eq./10 mL).

<https://doi.org/10.1371/journal.pone.0195738.t004>

Discussion

New strategies for Chagas disease diagnosis have been proposed with the introduction of commercially available tests. Their implementation should bring about a considerable improvement in the diagnosis of *T. cruzi* infection and lead to the standardization of protocols between




Fig 1. Degree of agreement between protocols A and H based on a Bland-Altman plot of the nine samples with quantifiable results in both methods. Each sample is represented by plotting the mean of the measurements obtained in protocols A and H on the x-axis and the difference of the same two values on the y-axis. SD: standard deviation. Log₁₀ par. eq./10 mL: logarithmic values of parasite equivalents in 10 mL of blood.

<https://doi.org/10.1371/journal.pone.0195738.g001>

Table 5. Cohen's kappa coefficient (K) results for the comparison between protocols with the sample treatment as the only variable. K values are shown including and excluding inhibited samples.

Protocols	N	K value (95% CI)	Positive samples	Inhibited samples	N excluding inhibition	Discordant samples excluding inhibition	K value excluding inhibited samples (95% CI)
A-E	25	0.92 (0.77–1)	11	0	25	1	0.92 (0.77–1)
B-F	25	0.76 (0.5–1)	10	0	25	3	0.76 (0.5–1)
C-G	25	0.45 (0.25–0.65)	5	8	17	1	0.87 (0.61–1)
D-H	25	0.55 (0.32–1)	7	6	19	1	0.89 (0.68–1)

N: number of samples assessed in parallel by both protocols compared. Positive samples: number of samples with a positive result for both protocols. Inhibited samples: number of samples with an invalid result for the internal amplification control (IAC) in at least one of the two protocols compared. CI: confidence interval.

<https://doi.org/10.1371/journal.pone.0195738.t005>

laboratories. Regarding molecular tests, to our knowledge only two Real-Time PCR assays (TCRUZIDNA.CE and RealCycler CHAG) are commercially available for the routine laboratory detection of *T. cruzi* DNA in clinical samples, both developed in Europe. Seiringer et al. [39] have recently assessed TCRUZIDNA.CE, which showed high sensitivity and specificity when compared with widely used PCR and Real-Time PCR strategies [26,46–48]. The study, performed in Italy, evaluated the PCR amplification but not the sample treatment or the extraction method. On the other hand, the RealCycler CHAG has not yet been assessed by comparison with other PCR assays, as recommended [30]. Thus, the aim of the work presented here was to assess this Real-Time PCR system taking into account the sample treatment and DNA extraction method, and is therefore the first study to evaluate the overall process including these three variables.

First of all, we created a panel of eight different protocols (Table 1) through the combination of three factors (sample treatments, DNA extraction methods, and Real-Time PCR procedures) and compared each one with the protocol considered the reference standard (protocol A in Table 1). All combinations reached a high concordance level except when the standard was compared with protocols starting from GEB and using the DNA extraction method based on magnetic particles (protocols C and D), which showed a high level of inhibition (Tables 1 and 2). A possible explanation is that the magnetic particle DNA isolation system was unable

Table 6. Cohen's kappa coefficient (K) results for the comparison between protocols with the DNA extraction system as the only variable. K values are shown including and excluding inhibited samples.

Protocols	N	K value (95% CI)	Positive samples	Inhibited samples	N excluding inhibition	Discordant samples excluding inhibition	K value excluding inhibited samples (95% CI)
A-C	62	0.17 (0.08–0.26)	8	35	27	8	0.45 (0.18–0.72)
B-D	62	0.46 (0.31–0.61)	23	17	45	4	0.82 (0.65–0.99)
E-G	25	0.92 (0.77–1)	11	0	25	1	0.92 (0.77–1)
F-H	25	0.76 (0.5–1)	10	0	25	3	0.76 (0.5–1)

N: number of samples assessed in parallel by both protocols compared. Positive samples: number of samples with a positive result for both protocols. Inhibited samples: number of samples with an invalid result for the internal amplification control (IAC) in at least one of the two protocols compared. CI: confidence interval.

<https://doi.org/10.1371/journal.pone.0195738.t006>

Table 7. Cohen's kappa coefficient (K) results for the comparison between protocols with the Real-Time PCR procedure as the only variable. K values are shown including and excluding inhibited samples.

Protocols	N	K value (95% CI)	Positive samples	Inhibited samples	N excluding inhibition	Discordant samples excluding inhibition	K value excluding inhibited samples (95% CI)
A-B	123	0.79 (0.68–0.9)	59	0	123	13 ^a	0.79 (0.68–0.9)
C-D	62	0.44 (0.29–0.6)	8	35	27	4	0.68 (0.41–0.96)
E-F	25	0.84 (0.62–1)	10	0	25	2	0.84 (0.62–1)
G-H	64	0.92 (0.81–1)	15	2	62	0	1

^aNone of these samples were positive in all three replicates.

N: number of samples assessed in parallel by both protocols compared. Positive samples: number of samples with a positive result for both protocols. Inhibited samples: number of samples with an invalid result for the internal amplification control (IAC) in at least one of the two protocols compared. CI: confidence interval.

<https://doi.org/10.1371/journal.pone.0195738.t007>

to completely eliminate the guanidine hydrochloride during the extraction process and the residual solution in the extracted DNA inhibited the PCR [35,49]. An alternative explanation is an excess of guanidine hydrochloride, given that some buffers of the kit reagent cartridges already contain guanidine salts. Indeed, all extracted DNA from inhibited samples showed a yellowish coloration, which was an early indication that the process had not worked properly. Moreover, it was noted that inhibition was related to the time elapsed between the sample treatment with guanidine and the DNA extraction: all inhibited samples were processed at least six months before being mixed with guanidine. In contrast, no inhibition was observed when DNA extraction occurred within the first week after the guanidine treatment. In the EB samples practically no inhibition occurred; the two exceptional samples (out of 64) had undergone magnetic particle DNA extraction (protocol G), and there was probably a residual pool of magnetic particles in the extracted DNA [50].

In this study, the RNase P human gene was used as an IAC to detect inhibition of the *in house* Real-Time PCR, as previously described by Pirón et al. [36]. However, we observed that patients with leukopenia or neutropenia showed later Ct values in the Real-Time PCR than immunocompetent individuals, rather than inhibition, as a consequence of their immunological condition. This should therefore be born in mind when hematological patients are included in studies of this kind. In fact, Duffy et al. [40] suggested the use of a heterologous extrinsic IAC, which has subsequently been taken up in other studies [42,51,52]. Likewise, the RealCycler assay includes this kind of control among the reagents supplied with the kit.

Once the inhibited samples were excluded from the analysis, discordant results appeared with regard to the extraction method and the number of replicates amplified in the Real-Time PCR. When comparing the reference protocol with those using GEB and magnetic particle DNA extraction (Table 2), the results suggest that silica column-based DNA extraction works more efficiently than the magnetic particle system with GEB samples. In contrast, the two DNA extraction methodologies presented similar results with EB samples (Table 6). Taking into account the number of replicates amplified in the Real-Time PCR, protocols using the RealCycler Real-Time PCR included only one reaction of amplification per sample instead of three, thus reducing the possibility of obtaining a positive result (Table 3). An increase in the number of replicates in these protocols would probably reduce conflicting results. Otherwise there were no important differences in results between protocols when comparing Real-Time PCRs (Table 7), which could be expected since both target the parasite SatDNA and probably use the same primer and probe sequences. However, the commercial formulation of the RealCycler CHAG assay is not known.

The results obtained highlight the importance of fine-tuning the overall process, including sample treatment, DNA extraction method and Real-Time PCR, for the successful diagnosis of *T. cruzi* infection. The reference standard protocol (A) and the one used for routine diagnosis in our hospital (H) were in perfect concordance, but the results were less convincing when factor variations were introduced (Table 2).

The quantitative analysis was only performed for protocols A and H because their comparison was the only one to show perfect agreement and no inhibition. The 15 samples positive with both protocols belonged to chronic patients, which probably explains why five samples had very low parasitic loads (samples S1, S2, S10, S11, and S14 in Table 4). The number of samples in our study was a limiting factor for the analysis of the results. According to Bland and Altman [44], some lack of agreement between different methods of measurement is inevitable, therefore what is important is the amount by which methods disagree. Based on this assumption, a Bland-Altman plot was constructed (Fig 1). As can be seen from the plot, the bias is not statistically significant, so protocols A and H were determined as equivalent and interchangeable [53].

Conclusions

The reference protocol (A) and the one used for routine molecular diagnosis of *T. cruzi* DNA in our hospital (H) were in perfect concordance and both tests can be used interchangeably. Therefore, protocol H (EB samples, EZ1 Virus Mini Kit v2.0 and the RealCycler CHAG) is a good option for the routine diagnosis of *T. cruzi* infection. The combination of these two assays allows results to be obtained in just over two hours, in a straightforward way and with minimum handling, which indicates its suitability for incorporation in hospital-associated laboratories. When variations in protocol factors were applied, the results were less convincing, which highlights that the overall process needs to be fine-tuned to obtain good results. The large number of inhibited GEB samples detected indicates they are not suitable for the DNA extraction method based on magnetic particles, at least when samples are not processed immediately.

Acknowledgments

The authors would like to thank the Werfen Group for the supply of material required for the study and Drs. Montserrat Portús, Ferran Navarro and Julio Alonso-Padilla for revising the manuscript.

Author Contributions

Conceptualization: Alba Abras, Cristina Ballart, Montserrat Gállego, Carmen Muñoz.

Data curation: Alba Abras, Cristina Ballart.

Formal analysis: Alba Abras, Cristina Ballart.

Funding acquisition: Joaquim Gascón, Montserrat Gállego, Carmen Muñoz.

Investigation: Alba Abras, Cristina Ballart.

Methodology: Alba Abras, Cristina Ballart, Montserrat Gállego, Carmen Muñoz.

Project administration: Montserrat Gállego, Carmen Muñoz.

Resources: Teresa Llovet, Carme Roig, Cristina Gutiérrez, Silvia Tebar, Pere Berenguer, María-Jesús Pinazo, Elizabeth Posada, Joaquim Gascón, Alejandro G. Schijman, Montserrat Gállego, Carmen Muñoz.

Software: Alba Abras, Cristina Ballart.

Supervision: Montserrat Gállego, Carmen Muñoz.

Validation: Alba Abras, Cristina Ballart.

Visualization: Alba Abras, Cristina Ballart, Montserrat Gállego, Carmen Muñoz.

Writing – original draft: Alba Abras, Cristina Ballart, Montserrat Gállego, Carmen Muñoz.

Writing – review & editing: Alba Abras, Cristina Ballart, María-Jesús Pinazo, Joaquim Gascón, Alejandro G. Schijman, Montserrat Gállego, Carmen Muñoz.

References

1. World Health Organization. Chagas disease in Latin America: an epidemiological update based on 2010 estimates. *Wkly Epidemiol Rec.* 2015; 90(6):33–44. PMID: [25671846](#)
2. Albajar-Viñas P, Jannin J. The hidden Chagas disease burden in Europe. *Euro Surveill.* 2011; 16(38): pii = 19975.
3. Gascón J, Bern C, Pinazo MJ. Chagas disease in Spain, the United States and other non-endemic countries. *Acta Trop.* 2010; 115(1–2):22–7. <https://doi.org/10.1016/j.actatropica.2009.07.019> PMID: [19646412](#)
4. Abad-Franch F, Lima MM, Sarquis O, Gurgel-Gonçalves R, Sánchez-Martín M, Calzada J, et al. On palms, bugs, and Chagas disease in the Americas. *Acta Trop.* 2015; 151:126–41. <https://doi.org/10.1016/j.actatropica.2015.07.005> PMID: [26196330](#)
5. Pérez-Molina JA, Molina I. Chagas disease. *Lancet.* 2017; S0140-6736(17):31612–4. [https://doi.org/10.1016/S0140-6736\(17\)31612-4](https://doi.org/10.1016/S0140-6736(17)31612-4)
6. Rassi A Jr, Rassi A, Marin-Neto JA. Chagas disease. *Lancet.* 2010; 375(9723):1388–402. [https://doi.org/10.1016/S0140-6736\(10\)60061-X](https://doi.org/10.1016/S0140-6736(10)60061-X) PMID: [20399979](#)
7. Prata A. Clinical and epidemiological aspects of Chagas disease. *Lancet Infect Dis.* 2001; 1(2):92–100. [https://doi.org/10.1016/S1473-3099\(01\)00065-2](https://doi.org/10.1016/S1473-3099(01)00065-2) PMID: [11871482](#)
8. Bern C. Chagas' disease. *N Engl J Med.* 2015; 373(19):1882. <https://doi.org/10.1056/NEJMc1510996> PMID: [26535522](#)
9. Rassi A Jr, Rassi A, Marcondes de Rezende J. American trypanosomiasis (Chagas disease). *Infect Dis Clin N Am.* 2012; 26(2):275–91. <https://doi.org/10.1016/j.idc.2012.03.002> PMID: [22632639](#)
10. Imaz-Iglesia I, García-San Miguel L, Ayala-Morillas LE, García-Pérez L, González-Enríquez J, Blasco-Hernández T, et al. Economic evaluation of Chagas disease screening in Spain. *Acta Trop.* 2015; 148:77–88. <https://doi.org/10.1016/j.actatropica.2015.04.014> PMID: [25917718](#)
11. Requena-Méndez A, Bussion S, Aldasoro E, Jackson Y, Angheben A, Moore D, et al. Cost-effectiveness of Chagas disease screening in Latin American migrants at primary health-care centres in Europe: a Markov model analysis. *Lancet Glob Health.* 2017; 5(4):e439–47. [https://doi.org/10.1016/S2214-109X\(17\)30073-6](https://doi.org/10.1016/S2214-109X(17)30073-6) PMID: [28256340](#)
12. Sicuri E, Muñoz J, Pinazo MJ, Posada E, Sánchez J, Alonso PL, et al. Economic evaluation of Chagas disease screening of pregnant Latin American women and of their infants in a non endemic area. *Acta Trop.* 2011; 118(2):110–7. <https://doi.org/10.1016/j.actatropica.2011.02.012> PMID: [21396345](#)
13. Picado A, Angheben A, Marchiol A, Alarcón de Noya B, Flevaud L, Pinazo MJ, et al. Development of diagnostic for Chagas disease: where should we put our limited resources? *PLoS Negl Trop Dis.* 2017; 11(1):e0005148. <https://doi.org/10.1371/journal.pntd.0005148> PMID: [28056025](#)
14. Abras A, Gállego M, Llovet T, Tebar S, Herrero M, Berenguer P, et al. Serological diagnosis of chronic Chagas disease: is it time for a change? *J Clin Microbiol.* 2016; 54(6):1566–72. <https://doi.org/10.1128/JCM.00142-16> PMID: [27053668](#)
15. World Health Organization. Control of Chagas disease: second report of the WHO expert committee. *World Health Organ Tech Rep Ser* 2002; 905:1–109.
16. Egúez KE, Alonso-Padilla J, Terán C, Chipana Z, García W, Torrico F, et al. Rapid diagnostic tests duo as alternative to conventional serological assays for conclusive Chagas disease diagnosis. *PLoS Negl Trop Dis.* 2017; 11(4):e0005501. <https://doi.org/10.1371/journal.pntd.0005501> PMID: [28369081](#)
17. Flores-Chávez M, Cruz I, Rodríguez M, Nieto J, Franco E, Gárate T, et al. Comparación de técnicas serológicas convencionales y no convencionales para el diagnóstico de la enfermedad de Chagas importada en España. *Enferm Infecc Microbiol Clin.* 2010; 28(5):284–93. <https://doi.org/10.1016/j.eimc.2009.07.005> PMID: [19962790](#)

18. Otani MM, Vinelli E, Kirchhoff LV, del Pozo A, Sands A, Vercauteren G, et al. WHO comparative evaluation of serologic assays for Chagas disease. *Transfusion*. 2009; 49(6):1076–82. <https://doi.org/10.1111/j.1537-2995.2009.02107.x> PMID: 19290995
19. Santos FLN, de Souza WV, Barros MDS, Nakazawa M, Krieger MA, Gomes YDM. Chronic Chagas disease: a comparative study performance of commercial enzyme immunoassay tests. *Am J Trop Med Hyg*. 2016; 94(5):1034–9. <https://doi.org/10.4269/ajtmh.15-0820> PMID: 26976886
20. Viotti R, Alarcón de Noya B, Arújo-Jorge T, Grijalva MJ, Guhl F, López MC, et al. Towards a paradigm shift in the treatment of chronic Chagas disease. *Antimicrob Agents Chemother*. 2014; 58(2):635–9. <https://doi.org/10.1128/AAC.01662-13> PMID: 24247135
21. Caballero ZC, Sousa OE, Marques WP, Sáez-Alquezar A, Umezawa ES. Evaluation of serological tests to identify *Trypanosoma cruzi* infection in humans and determine cross-reactivity with *Trypanosoma rangeli* and *Leishmania* spp. *Clin Vaccine Immunol*. 2007; 14(8):1045–9. <https://doi.org/10.1128/CI.00127-07> PMID: 17522327
22. Abras A, Muñoz C, Ballart C, Berenguer P, Llovet T, Herrero M, et al. Towards a new strategy for diagnosis of congenital *Trypanosoma cruzi* infection. *J Clin Microbiol*. 2017; 55(5):1396–407. <https://doi.org/10.1128/JCM.02248-16> PMID: 28202792
23. Carlier Y, Torrico F, Sosa-Estani S, Russomando G, Luquetti A, Freilij H, et al. Congenital Chagas disease: recommendations for diagnosis, treatment and control of newborns, siblings and pregnant women. *PLoS Negl Trop Dis*. 2011; 5(10):e1250. <https://doi.org/10.1371/journal.pntd.0001250> PMID: 22039554
24. Carlier Y, Sosa-Estani S, Luquetti AO, Buekens P. Congenital Chagas disease: an update. *Mem Inst Oswaldo Cruz*. 2015; 110(3):363–8. <https://doi.org/10.1590/0074-02760140405> PMID: 25760448
25. Balouz V, Agüero F, Buscaglia CA. Chagas disease diagnostic applications: present knowledge and future steps. *Adv Parasitol*. 2017; 97:1–45. <https://doi.org/10.1016/bs.apar.2016.10.001> PMID: 28325368
26. Schijman AG, Bisio M, Orellana L, Sued M, Duffy T, Mejía Jaramillo AM, et al. International study to evaluate PCR methods for detection of *Trypanosoma cruzi* DNA in blood samples from Chagas disease patients. *PLoS Negl Trop Dis*. 2011; 5(1):e931. <https://doi.org/10.1371/journal.pntd.0000931> PMID: 21264349
27. Bua J, Volta BJ, Perrone AE, Scollo K, Velázquez EB, Ruiz AM, et al. How to improve the early diagnosis of *Trypanosoma cruzi* infection: relationship between validated conventional diagnosis and quantitative DNA amplification in congenitally infected children. *PLoS Negl Trop Dis*. 2013; 7(10):e2476. <https://doi.org/10.1371/journal.pntd.0002476> PMID: 24147166
28. Burgos JM, Althech J, Petrucelli N, Bisio M, Levin MJ, Freilij H, et al. Molecular diagnosis and treatment monitoring of congenital transmission of *Trypanosoma cruzi* to twins of a triplet delivery. *Diagn Microbiol Infect Dis*. 2009; 65(1):58–61. <https://doi.org/10.1016/j.diagmicrobio.2009.04.010> PMID: 19679236
29. Rivero R, Bisio M, Velázquez EB, Esteva MI, Scollo K, González NL, et al. Rapid detection of *Trypanosoma cruzi* by colorimetric loop-mediated isothermal amplification (LAMP): A potential novel tool for the detection of congenital Chagas infection. *Diagn Microbiol Infect Dis*. 2017; 89(1):26–8. <https://doi.org/10.1016/j.diagmicrobio.2017.06.012> PMID: 28684052
30. Alonso-Padilla J, Gállego M, Schijman AG, Gascón J. Molecular diagnostics for Chagas disease: up to date and novel methodologies. *Expert Rev Mol Diagn*. 2017; 17(7):699–710. <https://doi.org/10.1080/14737159.2017.1338566> PMID: 28582629
31. Schijman AG. Molecular diagnosis of *Trypanosoma cruzi*. *Acta Trop*. 2018; pii: S0001-706X(17)30884-7. <https://doi.org/10.1016/j.actatropica.2018.02.019> PMID: 29476727
32. Ferrer E. Técnicas moleculares para el diagnóstico de la enfermedad de Chagas. *Saber*. 2015; 27(3):359–71.
33. Aebischer A, Beer M, Hoffmann B. Development and validation of rapid magnetic particle based extraction protocols. *Virol J*. 2014; 11:137. <https://doi.org/10.1186/1743-422X-11-137> PMID: 25086594
34. Berensmeier S. Magnetic particles for the separation and purification of nucleic acids. *Appl Microbiol Biotechnol*. 2006; 73(3):495–504. <https://doi.org/10.1007/s00253-006-0675-0> PMID: 17063328
35. Duffy T, Bisio M, Althech J, Burgos JM, Díez M, Levin MJ, et al. Accurate real-time PCR strategy for monitoring bloodstream parasitic loads in Chagas disease patients. *PLoS Negl Trop Dis*. 2009; 3(4):e419. <https://doi.org/10.1371/journal.pntd.0000419> PMID: 19381287
36. Pirón M, Fisa R, Casamitjana N, López-Chejade P, Puig L, Vergés M, et al. Development of a real-time PCR assay for *Trypanosoma cruzi* detection in blood samples. *Acta Trop*. 2007; 103(3):195–200. <https://doi.org/10.1016/j.actatropica.2007.05.019> PMID: 17662227
37. Elias MCQB, Vargas NS, Zingales B, Schenkman S. Organization of satellite DNA in the genome of *Trypanosoma cruzi*. *Mol Biochem Parasitol*. 2003; 129(1):1–9. PMID: 12798501

38. Elias MCQB, Vargas N, Tomazi L, Pedroso A, Zingales B, Schenkman S, et al. Comparative analysis of genomic sequences suggests that *Trypanosoma cruzi* CL Brener contains two sets of non-intercalated repeats of satellite DNA that corresponds to *T. cruzi* I and *T. cruzi* II types. *Mol Biochem Parasitol*. 2005; 140(2):221–7. <https://doi.org/10.1016/j.molbiopara.2004.12.016> PMID: 15760661
39. Seiringer P, Pritsch M, Flores-Chávez M, Marchisio E, Helfrich K, Mengele C, et al. Comparison of four PCR methods for efficient detection of *Trypanosoma cruzi* in routine diagnostics. *Diagn Microbiol Infect Dis*. 2017; 88(3):225–32. <https://doi.org/10.1016/j.diagmicrobio.2017.04.003> PMID: 28456430
40. Duffy T, Cura CI, Ramírez JC, Abate T, Cayo NM, Parrado R, et al. Analytical performance of a multiplex Real-Time PCR assay using TaqMan probes for quantification of *Trypanosoma cruzi* satellite DNA in blood samples. *PLoS Negl Trop Dis*. 2013; 7(1):e2000. <https://doi.org/10.1371/journal.pntd.0002000> PMID: 23350002
41. Burns MJ, Nixon GJ, Foy CA, Harris N. Standardisation of data from real-time quantitative PCR methods—evaluation of outliers and comparison of calibration curves. *BMC Biotechnol*. 2005; 5:31. <https://doi.org/10.1186/1472-6750-5-31> PMID: 16336641
42. Ramírez JC, Cura CI, da Cruz Moreira O, Lages-Silva E, Juiz N, Velázquez E, et al. Analytical validation of quantitative real-time PCR methods for quantification of *Trypanosoma cruzi* DNA in blood samples from Chagas disease patients. *J Mol Diagn*. 2015; 17(5):605–15. <https://doi.org/10.1016/j.jmoldx.2015.04.010> PMID: 26320872
43. Cohen J. A coefficient of agreement for nominal scales. *Educ Psychol Meas*. 1960; 20:37–46.
44. Bland JM, Altman DG. Measuring agreement in method comparison studies. *Stat Methods Med Res*. 1999; 8(2):135–60. <https://doi.org/10.1177/096228029900800204> PMID: 10501650
45. Spearman C. The proof and measurement of association between two things. *Am J Psychol*. 1904; 15:72–101.
46. Cummings KL, Tarleton RL. Rapid quantification of *Trypanosoma cruzi* in host tissue by real-time PCR. *Mol Biochem Parasitol*. 2003; 129(1):53–9. PMID: 12798506
47. Sturm NR, Degraeve W, Morel C, Simpson L. Sensitive detection of schizodeme classification of *Trypanosoma cruzi* cells by amplification of kinetoplast minicircle DNA sequences: use in diagnosis of Chagas disease. *Mol Biochem Parasitol*. 1989; 33(3):205–14. PMID: 2565018
48. Telleria J, Lafay B, Virreira M, Barnabé C, Tibayrenc M, Svoboda M. *Trypanosoma cruzi*: Sequence analysis of the variable region of kinetoplast minicircles. *Exp Parasitol*. 2006; 114(4):279–88. <https://doi.org/10.1016/j.exppara.2006.04.005> PMID: 16730709
49. Poma HR, Davies C, Gutiérrez Cacciabue D, Mora MC, Basombrío MA, Rajal VB. Comparación de la eficiencia de extracción de ácidos nucleicos utilizando distintos kits comerciales y empleando qPCR. Efecto de las sustancias inhibidoras. *Rev Argent Microbiol*. 2012; 44:144–9. PMID: 23102460
50. Dundas N, Leos NK, Mitui M, Revell P, Rogers BB. Comparison of automated nucleic acid extraction methods with manual extraction. *J Mol Diagn*. 2008; 10(4):311–6. <https://doi.org/10.2353/jmoldx.2008.070149> PMID: 18556770
51. Besuschio SA, Llano Murcia M, Benatar AF, Monnerat S, Cruz I, Picado A, et al. Analytical sensitivity and specificity of a loop-mediated isothermal amplification (LAMP) kit prototype for detection of *Trypanosoma cruzi* DNA in human blood samples. *PLoS Negl Trop Dis*. 2017; 11(7):e0005779. <https://doi.org/10.1371/journal.pntd.0005779> PMID: 28727723
52. Cura CI, Ramírez JC, Rodríguez M, Lopez-Albizu C, Irazu L, Scollo K, et al. Comparative study and analytical verification of PCR methods for the diagnosis of congenital Chagas disease. *J Mol Diagn*. 2017; 19(5):673–81. <https://doi.org/10.1016/j.jmoldx.2017.05.010> PMID: 28727985
53. Burd EM. Validation of laboratory-developed molecular assays for infectious diseases. *Clin Microbiol Rev*. 2010; 23(3):240–53. <https://doi.org/10.1128/CMR.00074-09> PMID: 20610823